



TURKISH ACCREDITATION AGENCY

ACCREDITATION CERTIFICATE

As a Calibration Laboratory

UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED
ŞİRKETİ

Central Address: OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye

is accredited in accordance with TS EN ISO/IEC 17025:2017 standard within the scope given in Annex following the assessment conducted by TURKAK.

Accreditation Number : AB-0012-K

Accreditation Date : 30.11.2004

Revision Date / Number : 15.04.2026 / 23

This certificate shall remain in force until 16.12.2029, subject to continuing compliance with the standard TS EN ISO/IEC 17025:2017, related regulations and requirements.

Gülden Banu Müderrisoğlu
Secretary General



Turkish Accreditation Agency (TURKAK) is a signatory to the European co-operation for Accreditation (EA) Multilateral Agreement (MLA) and International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Agreement (MRA) in the scope of ISO/IEC 17025.

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.


 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>		
	<p>Calibration Laboratory</p> <p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p> <p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>		

Calibration and Measurement Capability (CMC)


Temperature

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	0 °C	Ice Point	0,03 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory or in customer's site
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	-40 °C ≤ T ≤ 80 °C	Liquid Bath	0,08 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory or in customer's site
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	80 °C < T ≤ 250 °C	Liquid Bath	0,12 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory or in customer's site
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	-90 °C ≤ T ≤ 0 °C	Block Calibrator	0,10 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	-35 °C ≤ T ≤ 150 °C	Block Calibrator	0,12 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory or in customer's site
Resistance Thermometers • Platinum Resistance Thermometers (PRT) • Industrial Platinum Resistance Thermometers (PRT) • Thermistor • Transmitters	50 °C ≤ T ≤ 400 °C	Block Calibrator	0,20 °C	Calibration is performed according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 document. • Carried out in laboratory or in customer's site
Thermocouples Industrial Thermocouples	0 °C	Ice Point	0,10 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site • (except type B at negative temperatures)

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
	<p>Calibration Laboratory</p>			
<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>		<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>		
Thermocouples Industrial Thermocouples	$-40\text{ °C} \leq T \leq 80\text{ °C}$	Liquid Bath	0,16 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$80\text{ °C} < T \leq 250\text{ °C}$	Liquid Bath	0,33 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$-90\text{ °C} \leq T \leq 0\text{ °C}$	Block Calibrator	0,19 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$-35\text{ °C} \leq T \leq 150\text{ °C}$	Block Calibrator	0,25 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$50\text{ °C} \leq T \leq 400\text{ °C}$	Block Calibrator	0,50 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$400\text{ °C} < T \leq 600\text{ °C}$	Block Calibrator	1,5 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Thermocouples Industrial Thermocouples	$600\text{ °C} < T \leq 1100\text{ °C}$	Block Calibrator	2,2 °C	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site (except type B at negative temperatures)
Liquid-in-glass Thermometer	0 °C	Ice Point	0,03 °C	<p>For glass thermometers with 0.05 °C division and above calibration is performed using the comparison method.</p> <ul style="list-style-type: none"> Carried out in laboratory
Liquid-in-glass Thermometer	$-40\text{ °C} \leq T \leq 80\text{ °C}$	Liquid Bath	0,09 °C	<p>For glass thermometers with 0.05 °C division and above calibration is performed using the comparison method.</p> <ul style="list-style-type: none"> Carried out in laboratory

Accreditation Scope

 Calibration TS EN ISO/IEC 17025 AB-0012-K	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
	Calibration Laboratory Address : OŞTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr		

Liquid-in-glass Thermometer	$80\text{ }^{\circ}\text{C} < T \leq 250\text{ }^{\circ}\text{C}$	Liquid Bath	0,12 °C	For glass thermometers with 0.05 °C division and above calibration is performed using the comparison method. • Carried out in laboratory
Temperature Meters With Indicator Thermistor Resistance Transmitter	0 °C	Ice Point	0,03 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermistor Resistance Transmitter	$-40\text{ }^{\circ}\text{C} \leq T \leq 80\text{ }^{\circ}\text{C}$	Liquid Bath	0,08 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermistor Resistance Transmitter	$80\text{ }^{\circ}\text{C} < T \leq 250\text{ }^{\circ}\text{C}$	Liquid Bath	0,12 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermistor Resistance Transmitter	$-90\text{ }^{\circ}\text{C} \leq T \leq 0\text{ }^{\circ}\text{C}$	Block Calibrator	0,10 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory
Temperature Meters With Indicator Thermistor Resistance Transmitter	$-35\text{ }^{\circ}\text{C} \leq T \leq 150\text{ }^{\circ}\text{C}$	Block Calibrator	0,12 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermistor Resistance Transmitter	$50\text{ }^{\circ}\text{C} \leq T \leq 400\text{ }^{\circ}\text{C}$	Block Calibrator	0,20 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with the EN 60751 ve DKD R5-1 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	0 °C	Ice Point	0,26 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$-40\text{ }^{\circ}\text{C} \leq T \leq 80\text{ }^{\circ}\text{C}$	Liquid Bath	0,28 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$80\text{ }^{\circ}\text{C} < T \leq 250\text{ }^{\circ}\text{C}$	Liquid Bath	0,40 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>		
	<p>Calibration Laboratory</p>		
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>	<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$-90\text{ °C} \leq T \leq 0\text{ °C}$	Block Calibrator	0,28 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$-35\text{ °C} \leq T \leq 150\text{ °C}$	Block Calibrator	0,40 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$50\text{ °C} \leq T \leq 400\text{ °C}$	Block Calibrator	0,60 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$400\text{ °C} < T \leq 600\text{ °C}$	Block Calibrator	1,5 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Base Metal T,E,J,K,N Type) Transmitter	$600\text{ °C} < T \leq 1100\text{ °C}$	Block Calibrator	3,0 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Noble Metal R, S Type) Transmitter	0 °C	Ice Point	0,26 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Noble Metal R, S Type) Transmitter	$-40\text{ °C} \leq T \leq 80\text{ °C}$	Liquid Bath	0,26 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Noble Metal R, S Type) Transmitter	$80\text{ °C} < T \leq 250\text{ °C}$	Liquid Bath	0,30 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site
Temperature Meters With Indicator Thermocouple Sensor (Noble Metal R, S Type) Transmitter	$-50\text{ °C} \leq T \leq 0\text{ °C}$	Block Calibrator	0,26 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory
Temperature Meters With Indicator Thermocouple Sensor (Noble Metal R, S Type) Transmitter	$-35\text{ °C} \leq T \leq 150\text{ °C}$	Block Calibrator	0,40 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents. • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
	<p>Calibration Laboratory</p>			
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>		<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

<p>Temperature Meters With Indicator</p> <p>Thermocouple Sensor (Noble Metal R, S Type) Transmitter</p>	<p>$50\text{ °C} \leq T \leq 400\text{ °C}$</p>	<p>Block Calibrator</p>	<p>0,40 °C</p>	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Meters With Indicator</p> <p>Thermocouple Sensor (Noble Metal R, S Type) Transmitter</p>	<p>$400\text{ °C} < T \leq 600\text{ °C}$</p>	<p>Block Calibrator</p>	<p>1,5 °C</p>	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Meters With Indicator</p> <p>Thermocouple Sensor (Noble Metal R, S Type) Transmitter</p>	<p>$600\text{ °C} < T \leq 1100\text{ °C}$</p>	<p>Block Calibrator</p>	<p>2,2 °C</p>	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Meters With Indicator</p> <p>Thermocouple Sensor (Noble Metal B Type) Transmitter</p>	<p>$600\text{ °C} \leq T \leq 1100\text{ °C}$</p>	<p>Block Calibrator</p>	<p>2,2 °C</p>	<p>Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.8, EN 60584 and DKD R5-3 documents.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hygrometers</p> <p>Dew Point Meter Hygrometer Relative Humidity Meter (Capacitive, resistive, thermograph, mechanical, wet/dry chamber) Relative Humidity Meter (Datalogger) Relative Humidity Meter (Digital/Analog) Moisture Meter Humidity Transmitter</p>	<p>$10\text{ %rh} \leq RH \leq 95\text{ %rh}$</p>	<p>$15\text{ °C} \leq T \leq 35\text{ °C}$</p>	<p>2,0 %rh</p>	<p>Calibration is performed by using the comparison method with the reference Pt100 and temperature humidity meter in the humidity chamber according to the calibration procedure prepared in accordance with the DKD-R 5-8 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Hygrometers</p> <p>Hygrometer Relative Humidity Meter (Capacitive, resistive, thermograph, mechanical, wet/dry chamber) Relative Humidity Meter (Datalogger) Relative Humidity Meter (Digital/Analog) Moisture Meter Temperature Transmitter</p>	<p>$10\text{ °C} \leq T \leq 50\text{ °C}$</p>	<p>$30\text{ %rh} \leq RH \leq 60\text{ %rh}$</p>	<p>0,20 °C</p>	<p>Calibration is performed by using the comparison method with the reference Pt100 and temperature humidity meter in the humidity chamber according to the calibration procedure prepared in accordance with the DKD-R 5-8 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Hygrometers</p> <p>Dew Point Meter Hygrometer Relative Humidity Meter (Capacitive, resistive, thermograph, mechanical, wet/dry chamber) Relative Humidity Meter (Datalogger) Relative Humidity Meter (Digital/Analog) Moisture Meter Humidity Transmitter</p>	<p>$10\text{ %rh} \leq RH \leq 95\text{ %rh}$</p>	<p>$15\text{ °C} \leq T \leq 35\text{ °C}$</p>	<p>1,5 %rh</p>	<p>Calibration is performed by using the comparison method with the reference psychometric sensor in the humidity chamber according to the calibration procedure prepared in accordance with the DKD-R 5-8 document.</p> <ul style="list-style-type: none"> Carried out in laboratory

Accreditation Scope

 TÜRKAK  Calibration TS EN ISO/IEC 17025 AB-0012-K	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
	Calibration Laboratory		
Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye		Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr	

Hygrometers Dew Point Meter Hygrometer Relative Humidity Meter (Capacitive, resistive, thermograph, mechanical, wet/dry chamber) Relative Humidity Meter (Datalogger) Relative Humidity Meter (Digital/Analog) Moisture Meter Temperature Transmitter	$10\text{ }^{\circ}\text{C} \leq T \leq 50\text{ }^{\circ}\text{C}$	$30\text{ \%rh} \leq RH \leq 60\text{ \%rh}$	$0,17\text{ }^{\circ}\text{C}$	Calibration is performed by using the comparison method with the reference psychometric sensor in the humidity chamber according to the calibration procedure prepared in accordance with the DKD-R 5-8 document. <ul style="list-style-type: none"> Carried out in laboratory
Hygrometers Ambient Thermometer	$50\text{ }^{\circ}\text{C} < T \leq 80\text{ }^{\circ}\text{C}$	With Resistance Sensor thermometer	$0,60\text{ }^{\circ}\text{C}$	Calibration was performed according to the calibration procedure prepared in accordance with the DKD-R 5-8 document. <ul style="list-style-type: none"> Carried out in laboratory
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$600\text{ }^{\circ}\text{C} \leq T \leq 1820\text{ }^{\circ}\text{C}$	Type B	$0,49\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$600\text{ }^{\circ}\text{C} \leq T \leq 1820\text{ }^{\circ}\text{C}$	Type B	$0,45\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$0\text{ }^{\circ}\text{C} \leq T \leq 1767\text{ }^{\circ}\text{C}$	Type S	$0,49\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$0\text{ }^{\circ}\text{C} \leq T \leq 1767\text{ }^{\circ}\text{C}$	Type S	$0,45\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$0\text{ }^{\circ}\text{C} \leq T \leq 1767\text{ }^{\circ}\text{C}$	Type R	$0,47\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$0\text{ }^{\circ}\text{C} \leq T \leq 1767\text{ }^{\circ}\text{C}$	Type R	$0,44\text{ }^{\circ}\text{C}$	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope

 TÜRKAK  Calibration TS EN ISO/IEC 17025 AB-0012-K	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ		
	Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
Calibration Laboratory			
Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye		Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr	

Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$-200\text{ °C} \leq T \leq 1372\text{ °C}$	Type K	0,22 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$-200\text{ °C} \leq T \leq 1372\text{ °C}$	Type K	0,21 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$-200\text{ °C} \leq T \leq 1300\text{ °C}$	Type N	0,21 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$-200\text{ °C} \leq T \leq 1300\text{ °C}$	Type N	0,20 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$-210\text{ °C} \leq T \leq 1200\text{ °C}$	Type J	0,20 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$-210\text{ °C} \leq T \leq 1200\text{ °C}$	Type J	0,19 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$-250\text{ °C} \leq T \leq 400\text{ °C}$	Type T	0,19 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF	$-250\text{ °C} \leq T \leq 400\text{ °C}$	Type T	0,18 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON	$-250\text{ °C} \leq T \leq 1000\text{ °C}$	Type E	0,19 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>		
	<p>Calibration Laboratory</p>		
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>	<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF</p>	<p>$-250\text{ °C} \leq T \leq 1000\text{ °C}$</p>	<p>Type E</p>	<p>0,18 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p>	<p>$-200\text{ °C} \leq T \leq 850\text{ °C}$</p>	<p>Type PT100</p>	<p>0,05 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	<p>$600\text{ °C} \leq T \leq 1820\text{ °C}$</p>	<p>Type B</p>	<p>0,34 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	<p>$600\text{ °C} \leq T \leq 1820\text{ °C}$</p>	<p>Type B</p>	<p>0,28 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	<p>$0\text{ °C} \leq T \leq 1767\text{ °C}$</p>	<p>Type S</p>	<p>0,34 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	<p>$0\text{ °C} \leq T \leq 1767\text{ °C}$</p>	<p>Type S</p>	<p>0,27 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	<p>$0\text{ °C} \leq T \leq 1767\text{ °C}$</p>	<p>Type R</p>	<p>0,29 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	<p>$0\text{ °C} \leq T \leq 1767\text{ °C}$</p>	<p>Type R</p>	<p>0,24 °C</p>	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>		
	<p>Calibration Laboratory</p> <p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>		

<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	$-200\text{ °C} \leq T \leq 1372\text{ °C}$	Type K	0,15 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	$-200\text{ °C} \leq T \leq 1372\text{ °C}$	Type K	0,13 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	$-200\text{ °C} \leq T \leq 1300\text{ °C}$	Type N	0,13 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	$-200\text{ °C} \leq T \leq 1300\text{ °C}$	Type N	0,11 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	$-210\text{ °C} \leq T \leq 1200\text{ °C}$	Type J	0,11 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	$-210\text{ °C} \leq T \leq 1200\text{ °C}$	Type J	0,10 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction ON Temperature Calibrator</p>	$-250\text{ °C} \leq T \leq 400\text{ °C}$	Type T	0,10 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Temperature Indicators and Calibrators</p> <p>Temperature Indicators</p> <p>Cold Junction OFF Temperature Calibrator</p>	$-250\text{ °C} \leq T \leq 400\text{ °C}$	Type T	0,09 °C	<p>Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope

 Calibration TS EN ISO/IEC 17025 AB-0012-K	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026			
	Calibration Laboratory			
Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye		Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr		

Temperature Indicators and Calibrators Temperature Indicators Cold Junction ON Temperature Calibrator	-250 °C ≤ T ≤ 1000 °C	Type E	0,09 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. • Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Cold Junction OFF Temperature Calibrator	-250 °C ≤ T ≤ 1000 °C	Type E	0,09 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. • Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Temperature Calibrator	-200 °C ≤ T < 0 °C	Type PT100	0,02 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. • Carried out in laboratory or in customer's site
Temperature Indicators and Calibrators Temperature Indicators Temperature Calibrator	0 °C ≤ T ≤ 650 °C	Type PT100	0,05 °C	Calibration is performed using the electrical simulation method according to the calibration procedure prepared in accordance with the EURAMET cg.11 document. • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Cold Room (Deep Freezer etc.) Climate Chamber Sterilizer (Autoclave) Muffle Furnace	-90 °C ≤ T ≤ 80 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	0,70 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Cold Room (Deep Freezer etc.) Climate Chamber Sterilizer (Autoclave) Muffle Furnace	80 °C < T ≤ 140 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	1,0 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Cold Room (Deep Freezer etc.) Climate Chamber Sterilizer (Autoclave) Muffle Furnace	140 °C < T ≤ 200 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	1,4 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	200 °C < T ≤ 250 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	1,8 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
	<p>Calibration Laboratory</p>			
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>		<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

Controlled Volumes (Temperature Distribution) Oven Incubator Cold Room (Deep Freezer etc.) Climate Chamber Sterilizer (Autoclave)	$-90\text{ }^{\circ}\text{C} \leq T \leq 80\text{ }^{\circ}\text{C}$	Point temperature inside the cabin With reference thermocouple sensor thermometer	0,50 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	$80\text{ }^{\circ}\text{C} < T \leq 140\text{ }^{\circ}\text{C}$	Point temperature inside the cabin With reference thermocouple sensor thermometer	0,70 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 ve EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	$140\text{ }^{\circ}\text{C} < T \leq 200\text{ }^{\circ}\text{C}$	Point temperature inside the cabin With reference thermocouple sensor thermometer	0,90 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 ve EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	$200\text{ }^{\circ}\text{C} < T \leq 250\text{ }^{\circ}\text{C}$	Point temperature inside the cabin With reference thermocouple sensor thermometer	1,1 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 and EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Cold Room (Deep Freezer etc.) Climate Chamber Sterilizer (Autoclave)	$-90\text{ }^{\circ}\text{C} \leq T \leq 80\text{ }^{\circ}\text{C}$	Temperature distribution inside the cabin, With reference resistance sensor thermometer	0,50 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 ve EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	$80\text{ }^{\circ}\text{C} < T \leq 250\text{ }^{\circ}\text{C}$	Temperature distribution inside the cabin, With reference resistance sensor thermometer	0,70 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 ve EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Oven Incubator Climate Chamber Sterilizer (Autoclave)	$-90\text{ }^{\circ}\text{C} \leq T \leq 400\text{ }^{\circ}\text{C}$	Point temperature inside the cabin With reference resistance sensor thermometer	0,20 °C	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20, DKD-R 5-7 ve EN 60068 documents., • Carried out in laboratory or in customer's site
Controlled Volumes (Relative Humidity Distribution) Climate Chamber	$10\text{ } \% \text{rh} \leq RH \leq 90\text{ } \% \text{rh}$	Central Humidity ($15\text{ }^{\circ}\text{C} \leq T \leq 35\text{ }^{\circ}\text{C}$) Ambient	3,0 %rh	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20 documents., • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
	<p>Calibration Laboratory</p>			
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>		<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

Controlled Volumes (Relative Humidity Distribution) Climate Chamber	10 %rh ≤ RH ≤ 95 %rh	Central Humidity (5 °C ≤ T ≤ 85 °C) Ambient	2,0 %rh	Calibration is performed with the portable calibration system according to the calibration procedure prepared in accordance with EURAMET cg.20 documents., (With Psychometric Sensor) • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Autoclave, Steam Steriliser, Vacuum Etuve	30 °C ≤ T ≤ 140 °C	With Datalogger Temperature Measurement	0,3 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Autoclave, Steam Steriliser, Vacuum Etuve	30 °C ≤ T ≤ 140 °C	With Datalogger Temperature Measurement	0,042 bar	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Muffle Furnace	100 °C ≤ T ≤ 1400 °C	Axial Temperature Distribution	4,5 °C	With portable calibration system • Carried out in laboratory or in customer's site
Dry Block Calibrators	-90 °C ≤ T ≤ 400 °C	Homogeneity, Stability Indicator Deviation	0,4 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.13 documents. • Carried out in laboratory or in customer's site
Dry Block Calibrators	400 °C < T ≤ 1100 °C	Homogeneity, Stability Indicator Deviation	2,3 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.13 documents. • Carried out in laboratory or in customer's site
Dry Block Calibrators	1100 °C < T ≤ 1400 °C	Homogeneity, Stability Indicator Deviation	4,5 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with EURAMET cg.13 documents. • Carried out in laboratory or in customer's site
Industrial Radiation Thermometers Pyrometer IR Thermometer	20 °C ≤ T ≤ 250 °C	With IR Calibrator	2,1 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with ASTM E2847 documents. • Carried out in laboratory or in customer's site
Industrial Radiation Thermometers Pyrometer IR Thermometer	250 °C < T ≤ 400 °C	With IR Calibrator	3,2 °C	Calibration is performed using the comparison method according to the calibration procedure prepared in accordance with ASTM E2847 documents. • Carried out in laboratory or in customer's site
Industrial Radiation Thermometers Pyrometer IR Thermometer	50 °C ≤ T ≤ 600 °C	With IR Calibrator (Cavity)	3,2 °C	Calibration was performed according to the calibration procedure prepared in accordance with the ASTM E2847 document. Using the comparison method • Carried out in laboratory or in customer's site

Accreditation Scope

 Calibration TS EN ISO/IEC 17025 AB-0012-K	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
	Calibration Laboratory Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr		

Industrial Radiation Thermometers Pyrometer IR Thermometer	$600\text{ °C} < T \leq 1000\text{ °C}$	With IR Calibrator (Cavity)	6,0 °C	Calibration was performed according to the calibration procedure prepared in accordance with the ASTM E2847 document. Using the comparison method • Carried out in laboratory or in customer's site
Industrial Radiation Thermometers Pyrometer IR Thermometer	$1000\text{ °C} < T \leq 1200\text{ °C}$	With IR Calibrator (Cavity)	7,5 °C	Calibration was performed according to the calibration procedure prepared in accordance with the ASTM E2847 document. Using the comparison method • Carried out in laboratory or in customer's site
Other Thermometers Surface Thermometers	$20\text{ °C} \leq T \leq 50\text{ °C}$	With Surface Temperature Calibrator	0,5 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Other Thermometers Surface Thermometers	$50\text{ °C} < T \leq 100\text{ °C}$	With Surface Temperature Calibrator	0,8 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Other Thermometers Surface Thermometers	$100\text{ °C} < T \leq 150\text{ °C}$	With Surface Temperature Calibrator	1,1 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Other Thermometers Surface Thermometers	$150\text{ °C} < T \leq 200\text{ °C}$	With Surface Temperature Calibrator	1,4 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Other Thermometers Surface Thermometers	$200\text{ °C} < T \leq 250\text{ °C}$	With Surface Temperature Calibrator	1,7 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Hot Plate	$20\text{ °C} \leq T \leq 50\text{ °C}$	With surface temperature meter	1,0 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Hot Plate	$50\text{ °C} < T \leq 100\text{ °C}$	With surface temperature meter	1,3 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Hot Plate	$100\text{ °C} < T \leq 150\text{ °C}$	With surface temperature meter	1,5 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Hot Plate	$150\text{ °C} < T \leq 200\text{ °C}$	With surface temperature meter	2,0 °C	Calibration is performed using the comparison method. • Carried out in laboratory

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
	<p>Calibration Laboratory</p>			
	<p>Address : OSTİM OSB MAH. 1267/1 SK. No:5/ YENİMAHALLE/ANKARA Ankara / Türkiye</p>		<p>Phone : +90 312 385 5078 Fax : +90 312 385 5093 Email : kalibrasyon@umsankara.com.tr Website : www.umsankara.com.tr</p>	

Controlled Volumes (Temperature Distribution) Hot Plate	200 °C < T ≤ 250 °C	With surface temperature meter	2,5 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Hot Plate	250 °C < T ≤ 400 °C	With surface temperature meter	5,0 °C	Calibration is performed using the comparison method. • Carried out in laboratory
Controlled Volumes (Temperature Distribution) Liquid Bath	-90 °C ≤ T ≤ 80 °C	Temperature distribution inside the cabin, With reference resistance sensor thermometer	0,15 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	80 °C < T ≤ 250 °C	Temperature distribution inside the cabin, With reference resistance sensor thermometer	0,22 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	-90 °C ≤ T ≤ 80 °C	Point temperature inside the cabin With reference resistance sensor thermometer	0,05 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	80 °C < T ≤ 250 °C	Point temperature inside the cabin With reference resistance sensor thermometer	0,08 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	250 °C < T ≤ 400 °C	Point temperature inside the cabin With reference resistance sensor thermometer	0,12 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	-90 °C ≤ T ≤ 80 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	0,50 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	80 °C < T ≤ 140 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	0,60 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	140 °C < T ≤ 200 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	0,70 °C	With portable calibration system • Carried out in laboratory or in customer's site
Controlled Volumes (Temperature Distribution) Liquid Bath	200 °C < T ≤ 250 °C	Temperature distribution inside the cabin, With reference thermocouple sensor thermometer	0,80 °C	With portable calibration system • Carried out in laboratory or in customer's site

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Weighing Tools

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Balance	$m \leq 10$ kg	With E2 Class Mass	$1 \cdot 10^{-6}$	<p><i>m</i>: Balance Capacity Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.18 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Balance	$m \leq 40$ kg	With F1 Class Mass	$5,3 \cdot 10^{-6}$	<p><i>m</i>: Balance Capacity Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.18 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Balance	$m \leq 1000$ kg	With M1 Class Mass	$9,4 \cdot 10^{-5}$	<p><i>m</i>: Balance Capacity Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.18 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Balance	1000 kg < $m \leq 2000$ kg	Class M1 Mass and With Substitution Masses	$1,2 \cdot 10^{-4}$	<p><i>m</i>: Balance Capacity Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.18 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Non-automatic Weighing Devices Balance	1 mg $\leq m \leq 20$ mg	With E1 Class Mass	$2,5$ µg	<p><i>m</i>: Weighing Value Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.26 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Pressure

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Relative Pressure Piston Pressure Standard (Dead Weight Tester)	$0,1 \text{ MPa} \leq p \leq 4 \text{ MPa}$	Pneumatic	$3 \text{ Pa} + 1,07 \cdot 10^{-4} \cdot p$	p : Relative Pressure (Pa) Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.3 document. • Carried out in laboratory
Relative Pressure Piston Pressure Standard (Dead Weight Tester)	$0,1 \text{ MPa} \leq p \leq 6 \text{ MPa}$	Hydraulic	$5 \text{ Pa} + 5,45 \cdot 10^{-5} \cdot p$	p : Relative Pressure (Pa) Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.3 document. • Carried out in laboratory
Relative Pressure Piston Pressure Standard (Dead Weight Tester)	$6 \text{ MPa} < p \leq 120 \text{ MPa}$	Hydraulic	$122 \text{ Pa} + 8,79 \cdot 10^{-5} \cdot p$	p : Relative Pressure (Pa) Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.3 document. • Carried out in laboratory
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$1 \text{ bar} \leq p \leq 40 \text{ bar}$	Pneumatic (DWT)	$9,6 \cdot 10^{-5} \text{ bar} + 7,6 \cdot 10^{-5} \cdot p$	p : Relative Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$1 \text{ bar} \leq p \leq 60 \text{ bar}$	Hydraulic (DWT)	$1,2 \cdot 10^{-4} \text{ bar} + 2,4 \cdot 10^{-5} \cdot p$	p : Relative Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$60 \text{ bar} \leq p \leq 1200 \text{ bar}$	Hydraulic (DWT)	$2,2 \cdot 10^{-3} \text{ bar} + 5,4 \cdot 10^{-5} \cdot p$	p : Relative Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$-0,9 \text{ bar} \leq p \leq -0,04 \text{ bar}$	Pneumatic	$1,3 \cdot 10^{-4} \text{ bar} + 3,7 \cdot 10^{-4} \cdot p$	p : Relative Pressure [bar] Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site

Accreditation Scope




UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ


Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$0,015 \text{ bar} \leq p \leq 1 \text{ bar}$	Pneumatic	$1,3 \cdot 10^{-4} \text{ bar} + 3,7 \cdot 10^{-4} \cdot p$	p : Relative Pressure [bar] Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$-0,9 \text{ bar} \leq p \leq 2 \text{ bar}$	Pneumatic	0,63 mbar	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$2 \text{ bar} < p \leq 25 \text{ bar}$	Pneumatic	$1,8 \cdot 10^{-2} \text{ bar} + 4 \cdot 10^{-4} \cdot p$	p : Relative Pressure [bar] Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$10 \text{ bar} \leq p \leq 400 \text{ bar}$	Hydraulic	$6,3 \cdot 10^{-2} \text{ bar} + 5 \cdot 10^{-4} \cdot p$	p : Relative Pressure [bar] Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Analog Manometer Digital Manometer Pressure Calibrator Pressure Transducer Pressure Transmitter	$400 \text{ bar} < p \leq 700 \text{ bar}$	Hydraulic	$2,6 \cdot 10^{-1} \text{ bar} + 3 \cdot 10^{-4} \cdot p$	p : Relative Pressure [bar] Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Pressure Transducer Pressure Transmitter Difference Pressure Meter	$-2000 \text{ Pa} \leq p \leq -10 \text{ Pa}$	Pneumatic	3,1 Pa	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Pressure Transducer Pressure Transmitter Difference Pressure Meter	$10 \text{ Pa} \leq p \leq 2000 \text{ Pa}$	Pneumatic	3,1 Pa	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Relative Pressure Pressure Transducer Pressure Transmitter Difference Pressure Meter	$-10000 \text{ Pa} \leq p \leq -2001 \text{ Pa}$	Pneumatic	8,1 Pa	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site

Accreditation Scope

		UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
Relative Pressure Pressure Transducer Pressure Transmitter Difference Pressure Meter	$2001 \text{ Pa} \leq p \leq 10000 \text{ Pa}$	Pneumatic	8,1 Pa	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Absolute Pressure Analog Barometer Numeric Barometer	$750 \text{ hPa} \leq p \leq 1100 \text{ hPa}$	Pneumatic	0,4 hPa	Calibration is performed with Electromechanical calibrator according to the calibration procedure prepared in accordance with EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$1 \text{ bar} \leq p_{abs} \leq 41 \text{ bar}$	Pneumatic (DWT)	$9,6 \cdot 10^{-5} \text{ bar} + 7,5 \cdot 10^{-5} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$1 \text{ bar} \leq p_{abs} \leq 61 \text{ bar}$	Hydraulic (DWT)	$1,2 \cdot 10^{-4} \text{ bar} + 2,4 \cdot 10^{-5} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$61 \text{ bar} < p_{abs} \leq 1201 \text{ bar}$	Hydraulic (DWT)	$2,2 \cdot 10^{-3} \text{ bar} + 5,4 \cdot 10^{-5} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a piston manometer according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$0,015 \text{ bar} \leq p_{abs} \leq 3 \text{ bar}$	Pneumatic	0,6 mbar	p_{abs} : Absolute Pressure [bar] Calibration is performed with a electromechanical calibrator according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$2 \text{ bar} < p_{abs} \leq 26 \text{ bar}$	Pneumatic	$1,8 \cdot 10^{-2} \text{ bar} + 4 \cdot 10^{-4} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a electromechanical calibrator according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory or in customer's site
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	$10 \text{ bar} \leq p_{abs} \leq 401 \text{ bar}$	Hydraulic	$6,4 \cdot 10^{-2} \text{ bar} + 5 \cdot 10^{-4} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a electromechanical calibrator according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026			
Absolute Pressure Analog Manometer Digital Manometer Pressure Transducer Pressure Transmitter Pressure Calibrator	401 bar < $p_{abs} \leq 701$ bar	Hydraulic	$2,7 \cdot 10^{-1}$ bar + $3 \cdot 10^{-4} \cdot p_{abs}$	p_{abs} : Absolute Pressure [bar] Calibration is performed with a electromechanical calibrator according to the calibration procedure prepared in accordance with the EURAMET cg.17 document. <ul style="list-style-type: none">Carried out in laboratory or in customer's site

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Time and Frequency

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Frequency Frequency Meters Frequency Meters Optical Tachometer	$6 \text{ rpm} \leq \omega \leq 99,99 \text{ rpm}$	$r = 0,01 \text{ rpm}$	$0,013 \text{ rpm} + 5,6 \cdot 10^{-5} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution With the help of frequency applied to the optical LED with signal generator • Carried out in laboratory or in customer's site
Frequency Frequency Meters Frequency Meters Optical Tachometer	$99,99 \text{ rpm} < \omega \leq 999,9 \text{ rpm}$	$r = 0,1 \text{ rpm}$	$0,12 \text{ rpm} + 6,1 \cdot 10^{-5} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution With the help of frequency applied to the optical LED with signal generator • Carried out in laboratory or in customer's site
Frequency Frequency Meters Frequency Meters Optical Tachometer	$999,9 \text{ rpm} < \omega \leq 99999 \text{ rpm}$	$r = 1 \text{ rpm}$	$1,2 \text{ rpm} + 1,2 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution With the help of frequency applied to the optical LED with signal generator • Carried out in laboratory or in customer's site
Frequency Frequency Meters Frequency Meters Contact Tachometer	$250 \text{ rpm} \leq \omega \leq 999,9 \text{ rpm}$	$r = 0,1 \text{ rpm}$	$0,3 \text{ rpm} + 1,8 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution Calibration is performed using the comparison method with an optical tachometer. • Carried out in laboratory or in customer's site
Frequency Frequency Meters Frequency Meters Contact Tachometer	$999,9 \text{ rpm} < \omega \leq 1500 \text{ rpm}$	$r = 1 \text{ rpm}$	$2,0 \text{ rpm} + 1,2 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution Calibration is performed using the comparison method with an optical tachometer. • Carried out in laboratory or in customer's site
Frequency Frequency Sources Frequency Standard Frequency Generator (Centrifuge-Mixer Devices)	$6 \text{ rpm} \leq \omega \leq 99,99 \text{ rpm}$	$r = 0,01 \text{ rpm}$	$0,022 \text{ rpm} + 2,0 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution Calibration is performed using the direct measurement method with an optical tachometer. • Carried out in laboratory or in customer's site
Frequency Frequency Sources Frequency Standard Frequency Generator (Centrifuge-Mixer Devices)	$99,99 \text{ rpm} < \omega \leq 999,9 \text{ rpm}$	$r = 0,1 \text{ rpm}$	$0,22 \text{ rpm} + 2,0 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution Calibration is performed using the direct measurement method with an optical tachometer. • Carried out in laboratory or in customer's site
Frequency Frequency Sources Frequency Standard Frequency Generator (Centrifuge-Mixer Devices)	$999,9 \text{ rpm} < \omega \leq 99999 \text{ rpm}$	$r = 1 \text{ rpm}$	$1,9 \text{ rpm} + 1,6 \cdot 10^{-4} \cdot \omega$	ω : Measured Speed [rpm] r : Resolution Calibration is performed using the direct measurement method with an optical tachometer. • Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
---	---

<p>Time Range</p> <p>Interval Meters</p> <p>Time Difference Meter (Frequency Counter, Stopwatch, Timer)</p> <p>Time Interval Meters</p>	$1 \text{ s} \leq t \leq 7200 \text{ s}$	$r = 1 \text{ ms}$	$0,025 \text{ s} + 1,2 \cdot 10^{-5} \cdot t$	<p>r: Resolution t: Measured time interval [s] Calibration is performed using the direct measurement method with a reference stopwatch.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Time Range</p> <p>Interval Meters</p> <p>Time Difference Meter (Frequency Counter, Stopwatch, Timer)</p> <p>Time Interval Meters</p>	$7200 \text{ s} < t \leq 86400 \text{ s}$	$r = 1 \text{ ms}$	$0,002 \text{ s} + 1,5 \cdot 10^{-5} \cdot t$	<p>r: Resolution t: Measured time interval [s] Calibration is performed using the direct measurement method with a reference stopwatch.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p>	<p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	--	--

Calibration and Measurement Capability (CMC)

Dimensional Quantities

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
<p>Length Measuring Devices</p> <p>1-Dimension Progress Measuring Devices [Transducer, Actuator, Inductive Probe] (LVDT, PZT, ...)</p>	<p>$0 \text{ mm} \leq L \leq 100 \text{ mm}$</p>	<p>$r = 0,05 \text{ } \mu\text{m}$</p> <p>With 1-D Measuring Device</p>	<p>$(0,5 + 0,7.L) \text{ } \mu\text{m}$</p>	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Part 14.1 document. • Carried out in laboratory</p>
<p>Length Measuring Devices</p> <p>1-Dimension Progress Measuring Devices [Transducer, Actuator, Inductive Probe] (LVDT, PZT, ...)</p>	<p>$0 \text{ mm} \leq L \leq 100 \text{ mm}$</p>	<p>$r = 0,05 \text{ } \mu\text{m}$</p> <p>With Laser Interferometer</p>	<p>$(0,14 + 1,0.L) \text{ } \mu\text{m}$</p>	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Part 14.1 document. • Carried out in laboratory</p>
<p>Length Measuring Devices</p> <p>Gauge Block Comparators</p>	<p>$0,5 \text{ mm} \leq L \leq 100 \text{ mm}$</p>	<p>$r = 0,01 \text{ } \mu\text{m}$</p>	<p>33 nm</p>	<p>r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET document cg.02 document. • Carried out at the customer's site</p>
<p>Length Measuring Devices</p> <p>1-Dimension Measuring Device (Universal etc.)</p>	<p>$0 \text{ mm} \leq L \leq 300 \text{ mm}$</p>	<p>$r = 0,01 \text{ } \mu\text{m}$</p> <p>With Block Gauge</p>	<p>$(0,2 + 2.L) \text{ } \mu\text{m}$</p>	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 17.1 document. • Carried out at the customer's site</p>
<p>Length Measuring Devices</p> <p>Measure Clock Test Device (Gauge Calibrator, Comparator Calibrator etc.)</p>	<p>$0 \text{ mm} \leq L \leq 25 \text{ mm}$</p>	<p>$r = 0,001 \text{ mm}$</p> <p>With Electronic Measuring Probe</p>	<p>0,8 μm</p>	<p>r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.4 document. • Carried out in laboratory or in customer's site</p>
<p>Length Measuring Devices</p> <p>Measure Clock Test Device (Gauge Calibrator, Comparator Calibrator etc.)</p>	<p>$0 \text{ mm} \leq L \leq 100 \text{ mm}$</p>	<p>$r = 0,01 \text{ } \mu\text{m}$</p> <p>With Laser Interferometer</p>	<p>$(0,10 + 1,0.L) \text{ } \mu\text{m}$</p>	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 17.2 document. • Carried out in laboratory or in customer's site</p>
<p>Handheld Basic Measuring Devices</p> <p>Micrometer Head</p>	<p>$0 \text{ mm} \leq L \leq 25 \text{ mm}$</p>	<p>$r = 0,001 \text{ } \mu\text{m}$</p> <p>With 1-D Measuring Device</p>	<p>0,8 μm</p>	<p>r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.4 document. • Carried out in laboratory</p>

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
---	---	--	--	--

<p>Dimension Standards</p> <p>Gauge Block (short 0.5 mm - 100 mm)</p>	$0,5 \text{ mm} \leq L \leq 100 \text{ mm}$	<p>Centre Point Deviation with 1-D Measuring Instrument for Steel Gauge Blocks</p>	$(0,07 + 0,5 \cdot L) \text{ } \mu\text{m}$	<p><i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Gauge Block (short 0.5 mm - 100 mm)</p>	$0,5 \text{ mm} \leq L \leq 100 \text{ mm}$	<p>For Ceramic Gauge Blocks Centre with 1-D Measuring Device Point Deviation</p>	$(0,07 + 0,6 \cdot L) \text{ } \mu\text{m}$	<p><i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Gauge Block (short 0.5 mm - 100 mm)</p>	$0,5 \text{ mm} \leq L \leq 100 \text{ mm}$	<p>Centre Point Deviation with 1-D Measuring Instrument for Tungsten Carbide Gage Blocks</p>	$(0,07 + 1,3 \cdot L) \text{ } \mu\text{m}$	<p><i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Gauge Block (short 0.5 mm - 100 mm)</p>	$0,5 \text{ mm} \leq L \leq 100 \text{ mm}$	<p>Deviation Range (<i>v</i>)</p>	$0,05 \text{ } \mu\text{m}$	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Long Gauge Block (Length Bar) (125 mm- 1000 mm)</p>	$100 \text{ mm} < L \leq 300 \text{ mm}$	<p>Centre Point Deviation with 1-D Measuring Instrument for Steel Gauge Blocks</p>	$(0,15 + 1,5 \cdot L) \text{ } \mu\text{m}$	<p><i>L</i> : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Long Gauge Block (Length Bar) (125 mm- 1000 mm)</p>	$100 \text{ mm} < L \leq 300 \text{ mm}$	<p>For Ceramic Gauge Blocks Centre with 1-D Measuring Device Point Deviation</p>	$(0,15 + 1,7 \cdot L) \text{ } \mu\text{m}$	<p><i>L</i> : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Long Gauge Block (Length Bar) (125 mm- 1000 mm)</p>	$100 \text{ mm} < L \leq 300 \text{ mm}$	<p>Centre Point Deviation with 1-D Measuring Instrument for Tungsten Carbide Gage Blocks</p>	$(0,15 + 4 \cdot L) \text{ } \mu\text{m}$	<p><i>L</i> : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3650, VDI/VDE/DGQ 2618 Bölüm 3.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Micrometer Setting Bar [Flat, Screw]</p>	$25 \text{ mm} \leq L \leq 300 \text{ mm}$	<p>Centre Point Deviation with 1-D Measuring Device</p>	$(0,4 + 3 \cdot L) \text{ } \mu\text{m}$	<p><i>L</i> : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.4 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
<p>Dimension Standards</p> <p>Thickness gauge (Synthyl etc. (Feeler Gauge))</p>	$10 \text{ } \mu\text{m} \leq L \leq 2 \text{ mm}$	<p>3 Point Measurement with 1-D Measuring Device</p>	$0,7 \text{ } \mu\text{m}$	<p>Calibration is performed according to the calibration procedure prepared in accordance with the DIN 2275 document.</p> <ul style="list-style-type: none"> Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Line Standards Precision Line Scale Microscope Control Micrometry (Stage micrometry)	$0 \text{ mm} \leq L \leq 10 \text{ mm}$	Optical Reading Method With Reference Glass Ruler	$0,5 \text{ } \mu\text{m}$	Calibration is performed using the comparison method • Carried out in laboratory
Line Standards Precision Line Scale Microscope Control Micrometry (Stage micrometry)	$10 \text{ mm} < L \leq 300 \text{ mm}$	Optical Reading Method With Reference Glass Ruler	$(0,6 + 2,7 \cdot L) \text{ } \mu\text{m}$	L : Measured Value [m] Calibration is performed using the comparison method • Carried out in laboratory
Line Standards Precision Line Scale Microscope Control Micrometry (Stage micrometry)	$0 \text{ mm} \leq L \leq 300 \text{ mm}$	With Laser Interferometer	$(0,4 + 1,1 \cdot L) \text{ } \mu\text{m}$	L : Measured Value [m] Calibration is performed using the comparison method • Carried out in laboratory
Line Standards Steel Ruler, Workshop or Mechanical Business Scales	$0 \text{ mm} \leq L \leq 300 \text{ mm}$	Optical Reading Method With 2-D Measuring Device	$0,05 \text{ mm}$	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 8.2 document. • Carried out in laboratory
Line Standards Steel Ruler, Workshop or Mechanical Business Scales	$0 \text{ m} \leq L \leq 3 \text{ m}$	With Gauge Ruler	$(200 + 30 \cdot L) \text{ } \mu\text{m}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 8.2 document. • Carried out in laboratory
Line Standards Tape Meters (Land, Workshop, Pi) (Geodesic) Wire	$0 \text{ m} \leq L \leq 3 \text{ m}$	With Gauge Ruler	$(200 + 38 \cdot L) \text{ } \mu\text{m}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 8.2 document. • Carried out in laboratory
Line Standards Tape Meters (Land, Workshop, Pi) (Geodesic) Wire	$3 \text{ m} < L \leq 50 \text{ m}$	With Gauge Ruler	$(200 + 75 \cdot L) \text{ } \mu\text{m}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 8.2 document. • Carried out in laboratory
Diameter Standards Exterior Cylinder (Bumper Mastar (Ref, Pass- not pass, etc.) Piston, Pin (Screw Measuring Pins) Wire, Setting Gauge)	$0,2 \text{ mm} \leq D \leq 300 \text{ mm}$	Exterior with 1-D Measuring Device Diameter Measurement	$(0,7 + 2,4 \cdot D) \text{ } \mu\text{m}$	D : Measured Diameter [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.1 document. • Carried out in laboratory
Diameter Standards Exterior Cylinder (Bumper Mastar (Ref, Pass- not pass, etc.) Piston, Pin (Screw Measuring Pins) Wire, Setting Gauge)	$0,1 \text{ mm} \leq D \leq 20 \text{ mm}$	Exterior with 1-D Measuring Device Diameter Measurement	$0,4 \text{ } \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.2 document. • Carried out in laboratory

Accreditation Scope

UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ				
Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026				
				
Diameter Standards	$0,7 \text{ mm} \leq D \leq 200 \text{ mm}$	Inner with 1-D Measuring Device Diameter Measurement	$(0,6 + 2,5 \cdot D) \text{ } \mu\text{m}$	<i>D</i> : Measured Diameter [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.1 document. • Carried out in laboratory
Internal Cylinder (Ring Master (Ref, Passes-not pass etc.))				
Diameter Standards	$0,3 \text{ mm} \leq D \leq 100 \text{ mm}$	Diameter with 1-D Measuring Device Measurement	$(0,4 + 4,6 \cdot D) \text{ } \mu\text{m}$	<i>D</i> : Measured Diameter [m] Calibration is performed using the comparison method • Carried out in laboratory
Globe (Ball, Screw Measurement Probes (T-Probe))				
Angle Measuring Devices	$\alpha \leq 360^\circ$	$r = 30''$	0,4'	<i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 7.2 document. • Carried out in laboratory
(Bevel) Protractor				
Angle Measuring Devices	Base Length $L \leq 1000 \text{ mm}$	Parallelism	3,0 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 7.2 document. • Carried out in laboratory
(Bevel) Protractor				
Angle Measuring Devices	Base Length $L \leq 1000 \text{ mm}$	Straightness	2,0 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 7.2 document. • Carried out in laboratory
(Bevel) Protractor				
Angle Measuring Devices	$\alpha \leq 90^\circ$	$r = 0,001^\circ$	0,0012°	<i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the DIN 877, TS10832 document. • Carried out in laboratory
Clinometer				
Angle Measuring Devices	Base Length $L \leq 200 \text{ mm}$	$r = 0,001 \text{ mm/m}$	3,5 $\mu\text{m/m}$	<i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the DIN 877, TS10832 document. • Carried out in laboratory
Clinometer				
Angle Measuring Devices	Base Length $L \leq 200 \text{ mm}$	$r = 0,02 \text{ mm/m}$	6 $\mu\text{m/m}$	<i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the DIN 877, TS10832 document. • Carried out in laboratory
Electronic Level Meter Spirit Level				
Angle Measuring Devices	Base Length $L \leq 1000 \text{ mm}$	$r = 0,5 \text{ mm/m}$	0,3 mm/m	<i>r</i> : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the DIN 877, TS10832 document. • Carried out in laboratory
Electronic Level Meter Spirit Level				

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Angle Measuring Devices Electronic Level Meter Spirit Level	Base Length $L \leq 1000$ mm	Parallelism	3,0 μ m	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the DIN 877, TS10832 document. • Carried out in laboratory
Flatness Standards Optical Flat Optical Parallel (Parallel Slope)	$0 \text{ mm} \leq L \leq 100$ mm	Centre Thickness Measurement	0,7 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 6.1 document. • Carried out in laboratory
Flatness Standards Optical Flat Optical Parallel (Parallel Slope)	$0 \text{ mm} \leq L \leq 100$ mm	Parallelism Detection	0,08 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 6.1 document. • Carried out in laboratory
Flatness Standards Optical Flat Optical Parallel (Parallel Slope)	$60 \text{ mm} < D \leq 100$ mm	With Reference Glass Gauge Flatness Measurement	0,4 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 6.1 document. • Carried out in laboratory
Flatness Standards Optical Flat Optical Parallel (Parallel Slope)	$0 \text{ mm} \leq D \leq 60$ mm	With Gauge Comparator Flatness Measurement	0,2 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 6.1 document. • Carried out in laboratory
Flatness Standards Plate	$L \leq 8$ m	With Laser Interferometer Flatness Measurement $F = \pm 2$ mm	$(0,6 + 0,004.F) \mu$ m	L : Edge Length [m] F : Measured Flatness Error [μ m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 6.2 document. • Carried out at the customer's site
Screw Standards Flat Screw Plug Gauge	Section Department Diameter $D \leq 300$ mm	Step: 0.5 mm - 6 mm	2,5 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.8 document. • Carried out in laboratory
Screw Standards Straight Screw Ring Gauge	Section Department Diameter $D \leq 200$ mm	Step: 0.5 mm - 6 mm	2,7 μ m	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.9 document. • Carried out in laboratory
Handheld Basic Measuring Devices Screw Dental Combs	$a \leq 10$ mm (Step)	Step Measurement	10,0 μ m	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Handheld Basic Measuring Devices Screw Dental Combars	$a \leq 10$ mm (Step)	Angle Measurement	15'	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory
2 - Dimension 3-Dimension Measuring Devices Projection Device Profile Projection Measuring Microscope	X and Y Axes $L \leq 10$ mm	$r = 0.1 \mu\text{m}$ With Glass Ruler With Laser Interferometer	0,5 μm 0,2 μm	r : Resolution Calibration is performed using the comparison method. • Carried out at the customer's site
2 - Dimension 3-Dimension Measuring Devices Projection Device Profile Projection Measuring Microscope	X and Y Axes $10 \text{ mm} < L \leq 300$ mm	$r = 0.1 \mu\text{m}$ With Glass Ruler With Laser Interferometer	(0,7 + 4.L) μm (0,2 + 1.L) μm	L : Measured Value [m] r : Resolution Calibration is performed using the comparison method. • Carried out at the customer's site
2 - Dimension 3-Dimension Measuring Devices Projection Device Profile Projection Measuring Microscope	Angle Measurement $0^\circ \leq a \leq 360^\circ$	With Angle Gauge Blocks	0,3'	Calibration is performed using the comparison method. • Carried out at the customer's site
2 - Dimension 3-Dimension Measuring Devices Projection Device Profile Projection Measuring Microscope	Magnification Ratio	5X-100X	%1-%0,3	Calibration is performed using the comparison method. • Carried out at the customer's site
Handheld Basic Measuring Devices Outer Diameter Micrometry	$0 \text{ mm} \leq L \leq 2000$ mm	$r = 0,001$ mm	(1,3 + 25.L) μm	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.1 document. • Carried out in laboratory
Handheld Basic Measuring Devices Depth Micrometry	$0 \text{ mm} \leq L \leq 300$ mm	$r = 0,001$ mm	(1,0 + 10.L) μm	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.5 document. • Carried out in laboratory
Handheld Basic Measuring Devices Two Dotted Internal Diameter Micrometry	$0 \text{ mm} \leq L \leq 300$ mm	$r = 0,001$ mm	(1,0 + 10.L) μm	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.7 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Handheld Basic Measuring Devices Three-Point Inner Diameter Micrometry	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	$r = 0,001 \text{ mm}$	$(1,2 + 20.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10.8 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Caliper (External diameter, Internal diameter, Depth, Step measurements)	$0 \text{ mm} \leq L \leq 300 \text{ mm}$	$r = 0,005 \text{ mm}$	$(6 + 20.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 9.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Caliper (External diameter, Internal diameter, Depth, Step measurements)	$300 \text{ mm} < L \leq 3000 \text{ mm}$	$r = 0,01 \text{ mm}$	$(15 + 32.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 9.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Depth Calip	$0 \text{ mm} \leq L \leq 200 \text{ mm}$	$r = 0,005 \text{ mm}$	$(7 + 15.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 9.2 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Depth Calip	$200 \text{ mm} < L \leq 1000 \text{ mm}$	$r = 0,01 \text{ mm}$	$(13 + 25.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 9.2 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Height Gage	$0 \text{ mm} \leq L \leq 1000 \text{ mm}$	$r = 0,01 \text{ mm}$	$(13 + 25.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 9.3 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Measure Hours (Comparator)	$0 \text{ mm} \leq L \leq 10 \text{ mm}$	$r = 0,001 \text{ mm}$ Analogues	$1,0 \text{ } \mu\text{m}$	<p>r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 11.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory
Handheld Basic Measuring Devices Measure Hours (Comparator)	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	$r = 0,001 \text{ mm}$ Analogues	$(2 + 3.L) \text{ } \mu\text{m}$	<p>L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 11.1 document.</p> <ul style="list-style-type: none"> Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Handheld Basic Measuring Devices Measure Hours (Comparator)	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	$r = 0,5 \mu\text{m}$ Digital	$(0,7 + 6.L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 11.4 document. • Carried out in laboratory
Handheld Basic Measuring Devices Secretion Comparator (Precision Examiner)	$0 \text{ mm} \leq L \leq 2 \text{ mm}$	$r = 0,001 \text{ mm}$	$0,7 \mu\text{m}$	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 11.3 document. • Carried out in laboratory
Handheld Basic Measuring Devices Indicator	$0 \text{ mm} \leq L \leq 1 \text{ mm}$	$r = 0,001 \text{ mm}$	$0,6 \mu\text{m}$	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 11.2 document. • Carried out in laboratory
Handheld Basic Measuring Devices Fork Gauge (Interior, External)	$D \leq 250 \text{ mm}$	With 1-D Measuring Device	$(0,6 + 2,5.D) \mu\text{m}$	D : Measured Diameter [m] Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4,7 document. • Carried out in laboratory
Handheld Basic Measuring Devices Radius Gauges	$L \leq 50 \text{ mm}$	With 2-D Measuring Device	$10,0 \mu\text{m}$	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory
Handheld Basic Measuring Devices Hole Gauge (Bore Gauge etc.)	$0 \text{ mm} \leq L \leq 300 \text{ mm}$	$r = 0,001 \text{ mm}$	$(1,5 + 10.L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 13.2 document. • Carried out in laboratory
Handheld Basic Measuring Devices Thickness Meter (Interior, External, Passometers etc.)	$0 \text{ mm} \leq L \leq 50 \text{ mm}$	$r = 0,001 \text{ mm}$	$1,9 \mu\text{m}$	r : Resolution Calibration is performed using the comparison method with a parallel block gauge. • Carried out in laboratory
Handheld Basic Measuring Devices Thickness Meter (Interior, External, Passometers etc.)	$0 \text{ mm} \leq L \leq 2 \text{ mm}$	$r = 0,001 \text{ mm}$	$0,7 \mu\text{m}$	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 10,3 document. • Carried out in laboratory
Handheld Basic Measuring Devices Thickness Meter (Interior, External, Passometers etc.)	$0 \text{ mm} \leq L \leq 200 \text{ mm}$	$r = 0,001 \text{ mm}$	$3,0 \mu\text{m}$	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 12,1 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Handheld Basic Measuring Devices Thickness Meter (Interior, External, Passometers etc.)	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	$r = 0,001 \text{ mm}$	$3,0 \text{ } \mu\text{m}$	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 13,1 document. • Carried out in laboratory
Handheld Basic Measuring Devices Applikator	$L \leq 2 \text{ mm}$	Jeans Difference	$0,6 \text{ } \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the ASTM D 823 ve ASTM D4400 document. • Carried out in laboratory
Handheld Basic Measuring Devices Grindometer	$L \leq 100 \text{ } \mu\text{m}$	Jeans Difference	$0,6 \text{ } \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 1524 ve ASTM D1210 document. • Carried out in laboratory
Handheld Basic Measuring Devices Wet Film Thickness Gauge	$L \leq 5 \text{ mm}$	Jeans Difference	$2,0 \text{ } \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 2808 ve ASTM D4414 document. • Carried out in laboratory
Handheld Basic Measuring Devices Paint Adhesion Test Comb (Cross-Cut) EN ISO 2409	$L \leq 10 \text{ mm}$	Jeans Difference	$6,0 \text{ } \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 2409 document. • Carried out in laboratory
Handheld Basic Measuring Devices Paint Adhesion Test Comb (Cross-Cut) EN ISO 2409	$L \leq 10 \text{ mm}$	Angle Measurement	10'	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 2409 document. • Carried out in laboratory
Reference Materials [Sieve, Mesh] Aperture (Aperture)	Mesh Range $20 \text{ } \mu\text{m} \leq L \leq 5 \text{ mm}$	With 2-D Measuring Device	$(0,9 + 6,5L) \text{ } \mu\text{m}$	L : Measured Mesh Spacing [m] Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3310-1-2-3, TS 5458 ISO 5223 ve TS 9582 ISO 933 document. • Carried out in laboratory
Reference Materials [Sieve, Mesh] Aperture (Aperture)	Mesh Range $5 \text{ mm} < L \leq 125 \text{ mm}$	With the Digital Caliper	$(40 + 2,6L) \text{ } \mu\text{m}$	L : Measured Mesh Spacing [m] Calibration is performed according to the calibration procedure prepared in accordance with the ISO 3310-1-2-3, TS 5458 ISO 5223 ve TS 9582 ISO 933 document. • Carried out in laboratory
Reference Materials Concrete Sample Mold	$35 \text{ mm} \leq L \leq 300 \text{ mm}$	Parallelism	$0,03 \text{ mm}$	Calibration is performed according to the calibration procedure prepared in accordance with the TS 12390-1, TS 196-1, ASTM D1883-99, ASTM D1557 ve 698-07 document. • Carried out in laboratory

Accreditation Scope




UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Reference Materials Concrete Sample Mold	$35 \text{ mm} \leq L \leq 300 \text{ mm}$	Steepness	0,05 mm	Calibration is performed according to the calibration procedure prepared in accordance with the TS 12390-1, TS 196-1, ASTM D1883-99, ASTM D1557 ve 698-07 document. • Carried out in laboratory
Reference Materials Concrete Sample Mold	$35 \text{ mm} \leq L \leq 300 \text{ mm}$	Straightness	0,01 mm	Calibration is performed according to the calibration procedure prepared in accordance with the TS 12390-1, TS 196-1, ASTM D1883-99, ASTM D1557 ve 698-07 document. • Carried out in laboratory
Coating Thickness Coating Thickness Standard (Thickness Foils)	$10 \mu\text{m} \leq L \leq 50 \text{ mm}$	5 Point Measurement with 1-D Measuring Device	$(0,5 + 30.L) \mu\text{m}$	Calibration is performed according to the calibration procedure prepared in accordance with the TS 2311 EN ISO 2178 ve TS 2674 EN ISO 2360 document. • Carried out in laboratory
Coating Thickness Coating Thickness Measuring Device	$0 \text{ mm} \leq L \leq 2 \text{ mm}$	$r = 0,1 \mu\text{m}$ With Thickness Foil	0,8 μm	r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the TS 2311 EN ISO 2178 ve TS 2674 EN ISO 2360 document. • Carried out in laboratory
Length Measuring Devices 1-Dimension Measuring Device (Universal etc.)	$0 \text{ mm} \leq L \leq 1000 \text{ mm}$	$r = 0,01 \mu\text{m}$ With Laser Interferometer	$(0,1 + 0,5.L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 17.1 ve 17.2 document. • Carried out at the customer's site
2 - Dimension 3-Dimension Measuring Devices Three-Dimensional Measurement Device (CMM)	X, Y and Z Axes $0 \text{ m} \leq L \leq 10 \text{ m}$	With Laser Interferometer	$(0,6 + 1,8.L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 17.2 ve EN ISO 10360 document. • Carried out at the customer's site
Length Measuring Devices Single Axis Linear Measurement Systems (Tape Measure, Steel Ruler etc. Calibration Stands)	$0 \text{ m} \leq L \leq 10 \text{ m}$	$r = 0,001 \text{ mm}$ With Laser Interferometer	$(0,8 + 1,8.L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 17.2 document. • Carried out at the customer's site
Dimension Standards Gauge Block (short 0.5 mm - 100 mm)	$L < 0,5 \text{ mm}$	Centre with 1-D Measuring Device Point Deviation	0,18 μm For All Types	Calibration is performed using the comparison method • Carried out in laboratory

Accreditation Scope

		UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ		
Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026				
Diameter Standards Tapered Buffer Adjusting Gauge	$D \leq 200 \text{ mm}$	Exterior with 1-D Measuring Device Diameter Measurement	2,0 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.12 document. • Carried out in laboratory
Diameter Standards Tapered Ring Adjusting Gauge	$D \leq 200 \text{ mm}$	Inner with 1-D Measuring Device Diameter Measurement	2,0 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.12 document. • Carried out in laboratory
Screw Standards Tapered Screw Plug Gauge	Section Department Diameter $D \leq 300 \text{ mm}$	Step : 0.5 mm - 6 mm	3,0 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.12 document. • Carried out in laboratory
Screw Standards Tapered Screw Ring Gauge	Section Department Diameter $D \leq 200 \text{ mm}$	Step : 0.5 mm - 6 mm	2,7 μm	Calibration is performed according to the calibration procedure prepared in accordance with the VDI/VDE/DGQ 2618 Bölüm 4.12 document. • Carried out in laboratory
Handheld Basic Measuring Devices Ultrasonic Thickness Meter	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	$r = 0,001 \text{ mm}$	$(5,0 + 30 \cdot L) \mu\text{m}$	L : Measured Value [m] r : Resolution Calibration is performed using the comparison method • Carried out in laboratory
2 - Dimension 3-Dimension Measuring Devices Surface Roughness Measurement Device	$0,4 \mu\text{m} \leq Ra \leq 3 \mu\text{m}$	With roughness standard	$(23 + 33 \cdot Ra) \text{ nm}$	Ra : Measured Roughness Parameter, [μm] Calibration is performed according to the calibration procedure prepared in accordance with the DKD 4.2 Part 1 document. • Carried out in laboratory
2 - Dimension 3-Dimension Measuring Devices Surface Roughness Measurement Device	$1,6 \mu\text{m} \leq Rz \leq 11,5 \mu\text{m}$	With roughness standard	$(84 + 32 \cdot Rz) \text{ nm}$	Rz : Measured Roughness Parameter, [μm] Calibration is performed according to the calibration procedure prepared in accordance with the DKD 4.2 Part 1 document. • Carried out in laboratory
Handheld Basic Measuring Devices Welder's Caliper (Boden) Welding Visual Control Gauges (Welding Caliper)	$0 \text{ mm} \leq L \leq 50 \text{ mm}$	Height measurement	0,012 mm	Calibration is performed using the comparison method with a parallel block gauge. • Carried out in laboratory
Handheld Basic Measuring Devices Welding Visual Control Gauges (Welding Caliper) Interval Ruler	$0 \text{ mm} \leq L \leq 100 \text{ mm}$	Ruler measurement	0,02 mm	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory
Handheld Basic Measuring Devices Welding Visual Control Gauges (Welding Caliper)	$\alpha \leq 90^\circ$	Angle measurement	0,02°	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Handheld Basic Measuring Devices Chamfer Caliper	$0 \text{ mm} \leq L \leq 50 \text{ mm}$	Chamfer measurement	0,012 mm	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory
Long distance Laser meter	$0 \text{ mm} \leq L \leq 20 \text{ m}$	$r = 1 \text{ mm}$	$(2 + 0,16 \cdot L) \text{ mm}$	L : Measured Value [m] r : Resolution Calibration is performed using the comparison method with a reference laser meter. • Carried out in laboratory or in customer's site
Optical Standards Radius, Diameter Standards (Templates)	$D \leq 50 \text{ mm}$	With 2-D Measuring Device	2,3 μm	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory
Optical Standards Angle Templates	$\alpha \leq 360^\circ$	With 2-D Measuring Device	0,4'	Calibration is performed using the direct measurement method with the optical measurement method. • Carried out in laboratory

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p>	<p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	--	--

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Hardness

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
<p>Hardness Test Equipment Brinell Hardness Test Equipment</p>	<p>$100 \leq HB \leq 450$</p>	<p>HBW 10 / 3000</p>	<p>%0,6</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 ve ASTM E10 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment Brinell Hardness Test Equipment</p>	<p>$100 \leq HB \leq 450$</p>	<p>HBW 5 / 750</p>	<p>%0,9</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 ve ASTM E10 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment Brinell Hardness Test Equipment</p>	<p>$100 \leq HB \leq 450$</p>	<p>HBW 2,5 / 187,5</p>	<p>%1,2</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 ve ASTM E10 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment Brinell Hardness Test Equipment</p>	<p>$100 \leq HB \leq 450$</p>	<p>HBW 2,5 / 62,5</p>	<p>%0,8</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 ve ASTM E10 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Indenter Optical Track Measuring Equipment Inspection / Calibration</p>	<p>$L \leq 10$ mm</p>	<p>with stage micrometer</p>	<p>$1,5 \cdot 10^{-3} \cdot L$ (from 0.5 μm not being small provided that)</p>	<p>L : Measured Value [mm] Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 ve ASTM E10 document.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment Rockwell Hardness Test Equipment</p>	<p>HRB</p>		<p>1,0 HRB</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6508-2 ve ASTM E18 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Hardness Test Equipment Rockwell Hardness Test Equipment	HRC		0,5 HRC	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6508-2 ve ASTM E18 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Hardness Test Equipment Vickers Hardness Test Equipment	HV0.1		%5,4	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Hardness Test Equipment Vickers Hardness Test Equipment	HV0.3		%3,1	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Hardness Test Equipment Vickers Hardness Test Equipment	HV0.5		%2,7	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Hardness Test Equipment Vickers Hardness Test Equipment	HV1		%2,4	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
Hardness Test Equipment Vickers Hardness Test Equipment	HV2		%1,4	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
---	---	--	--	--

<p>Hardness Test Equipment</p> <p>Vickers Hardness Test Equipment</p>	<p>HV5</p>		<p>%1,0</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment</p> <p>Vickers Hardness Test Equipment</p>	<p>HV10</p>		<p>%1,0</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2, ASTM E384 ve ASTM E92 document.</p> <p>These values are based on indirect comparisons with hardness comparison plates. are the uncertainty values in the calibration. In direct calibration uncertainty values are also given.</p> <ul style="list-style-type: none"> Carried out at the customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment</p>	<p>$30 \leq HRC \leq 64$</p>	<p>Indirect calibration with hardness plates</p>	<p>1,8 HRC</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ISO 16859-2 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment</p>	<p>$550 \leq HLD \leq 850$</p>	<p>Indirect calibration with hardness plates</p>	<p>18 HLD</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ISO 16859-2 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment(With UCI method)</p>	<p>HV1</p>	<p>Indirect calibration with hardness plates</p>	<p>%2,4</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ASTM A 1038 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment(With UCI method)</p>	<p>HV2</p>	<p>Indirect calibration with hardness plates</p>	<p>%1,4</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ASTM A 1038 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment(With UCI method)</p>	<p>HV5</p>	<p>Indirect calibration with hardness plates</p>	<p>%1,0</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ASTM A 1038 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment(With UCI method)</p>	<p>HV10</p>	<p>Indirect calibration with hardness plates</p>	<p>%1,0</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ASTM A 1038 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
<p>Hardness Test Equipment</p> <p>Portable (Handheld) Hardness Test Equipment(With UCI method)</p>	<p>HRC</p>	<p>Indirect calibration with hardness plates</p>	<p>%0,5 HRC</p>	<p>Calibration is performed according to the calibration procedure prepared in accordance with the ASTM A 1038 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Hardness Test Equipment Shore hardness control plate	Shore A		0,6 Shore	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 48-4, ISO 48-9 ve ISO 868 document. • Carried out in laboratory or in customer's site
Hardness Test Equipment Shore hardness control plate	Shore D		0,6 Shore	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 48-4, ISO 48-9 ve ISO 868 document. • Carried out in laboratory or in customer's site
Hardness Test Equipment Shore Hardness Test Equipment	Shore A Shore D Shore AO Shore AM Shore B Shore C Shore DO Shore E Shore M Shore O Shore OO Shore OOO Shore OOO-S	Sinking Depth Force Diameter Radius Angle Hardness	2,0 µm %0,09 2,0 µm 2,0 µm 0,05° 0,5 Shore	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 48-4, ISO 48-9, ISO 868 ve ASTM 2240 document. • Carried out in laboratory
Hardness Test Equipment IRHD Hardness Test Equipment	Method N,H,L	Sinking Depth Force End Diameter Presser Foot and Hole Diameter Hardness	2,0 µm %0,09 2,0 µm 20,0 µm 0,5 IRHD	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 48-2 ve ISO 48-9 document. • Carried out at the customer's site
Hardness Test Equipment IRHD Hardness Test Equipment	Method M	Sinking Depth Force (8.3 mN) Force (145 mN-235 mN) End Diameter Presser Foot and Hole Diameter Hardness	0,4 µm %0,41 %0,08 1,0 µm 20,0 µm 0,5 IRHD	Calibration is performed according to the calibration procedure prepared in accordance with the ISO 48-2 ve ISO 48-9 document. • Carried out at the customer's site
Hardness Indenter Rockwell conical diamond (brale)	$\alpha = 120^\circ$ $R = 0,2 \text{ mm}$ $\alpha = 0,5^\circ$ $S = 0,4 \text{ mm}$	Angle Radius Axiality Straightness	0,05° 5,0 µm 0,05° 1,0 µm	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6508-2 ve ASTM E18 document. • Carried out in laboratory
Hardness Indenter Vickers diamond pyramide	$\alpha = 148,11^\circ / 136^\circ$ $\alpha = 0,5^\circ$	Pyramid Angle Axis Misalignment	0,05°	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6507-2 ve ASTM E92 document. • Carried out in laboratory
Hardness Indenter Brinell ve Rockwell hard steel ball (B-E-F-G-H-K-L-M-PR- S-T-V-W-X-Y)	$R = 1 - 1,5875 - 2,5 - 3,175 - 5 - 6,35 - 10 - 12,7 \text{ mm}$	Diameter of the Sphere	$(0,4 + 4,6.L) \text{ µm}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with the TS EN ISO 6506-2 - TS EN ISO 6508-2 ve ASTM E10 - ASTM E18 document. • Carried out in laboratory

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Volume

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Volumetric Apparatus Pipette (Piston)	1 μ L \leq V \leq 3 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,80	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	3 μ L < V \leq 5 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,60	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	5 μ L < V \leq 10 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,35	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	10 μ L < V \leq 50 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,30	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	50 μ L < V \leq 5 mL	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,20	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Apparatus Pipette (Piston)	5 mL < V ≤ 20 mL	Single channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,15	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	2 µL	Multi channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 2,6	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	2 µL < V ≤ 5 µL	Multi channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 1,0	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	5 µL < V ≤ 10 µL	Multi channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,60	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	10 µL < V ≤ 50 µL	Multi channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,40	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	50 µL < V ≤ 2 mL	Multi channel, Manual or motor-driven pipettes with piston movement (Type A and D1 pipettes, with digital and analogue display)	% 0,32	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Apparatus Pipette (Piston)	5 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type D2 pipettes, with digital and analogue display)	% 0,60	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	5 μ L < V \leq 20 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type D2 pipettes, with digital and analogue display)	% 0,40	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	20 μ L < V \leq 100 μ L	Single channel, Manual or motor-driven pipettes with piston movement (Type D2 pipettes, with digital and analogue display)	% 0,30	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Pipette (Piston)	100 μ L < V \leq 1 mL	Single channel, Manual or motor-driven pipettes with piston movement (Type D2 pipettes, with digital and analogue display)	% 0,25	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-2, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Syringes	1 μ L \leq V \leq 10 μ L	Piston motion by hand	% 1,5	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-9, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Syringes	10 μ L < V \leq 2 mL	Piston motion by hand	% 1,4	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-9, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Syringes	2 mL < V ≤ 200 mL	Piston motion by hand	% 1,2	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-9, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	V ≤ 1 mL	Piston motion by hand	% 0,17	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	1 mL < V ≤ 2 mL	Piston motion by hand	% 0,11	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	2 mL < V ≤ 10 mL	Piston motion by hand	% 0,08	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	10 mL < V ≤ 100 mL	Piston motion by hand	% 0,06	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	V ≤ 1 mL	Piston motion by motorized	% 0,17	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Burette (with Piston) -	1 mL < V ≤ 2 mL	Piston motion by motorized	% 0,12	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	2 mL < V ≤ 5 mL	Piston motion by motorized	% 0,08	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	5 mL < V ≤ 25 mL	Piston motion by motorized	% 0,06	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Burette (with Piston) -	25 mL < V ≤ 100 mL	Piston motion by motorized	% 0,05	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-3, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	10 µL	Single-stroke dispensers	% 0,55	The values given in the range column are nominal values. According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	10 µL < V ≤ 20 µL	Single-stroke dispensers	% 0,45	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Apparatus Dispenser	20 μ L < V \leq 50 μ L	Single-stroke dispensers	% 0,35	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	50 μ L < V \leq 100 μ L	Single-stroke dispensers	% 0,31	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	100 μ L < V \leq 200 μ L	Single-stroke dispensers	% 0,22	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	200 μ L < V \leq 500 μ L	Single-stroke dispensers	% 0,20	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	500 μ L < V \leq 200 mL	Single-stroke dispensers	% 0,14	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	1 μ L \leq V \leq 2 μ L	Multiple-delivery dispensers	% 1,3	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Apparatus Dispenser	$2 \mu\text{L} < V \leq 3 \mu\text{L}$	Multiple-delivery dispensers	% 0,80	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	$3 \mu\text{L} < V \leq 10 \mu\text{L}$	Multiple-delivery dispensers	% 0,60	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	$10 \mu\text{L} < V \leq 20 \mu\text{L}$	Multiple-delivery dispensers	% 0,50	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	$20 \mu\text{L} < V \leq 50 \mu\text{L}$	Multiple-delivery dispensers	% 0,30	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	$50 \mu\text{L} < V \leq 200 \mu\text{L}$	Multiple-delivery dispensers	% 0,25	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	$200 \mu\text{L} < V \leq 500 \mu\text{L}$	Multiple-delivery dispensers	% 0,23	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Apparatus Dispenser	500 μ L < V \leq 1 mL	Multiple-delivery dispensers	% 0,20	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	1 mL < V \leq 2 mL	Multiple-delivery dispensers	% 0,16	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	2 mL < V \leq 25 mL	Multiple-delivery dispensers	% 0,12	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Apparatus Dispenser	25 mL < V \leq 200 mL	Multiple-delivery dispensers	% 0,10	V: Nominal Volume According to the calibration procedure prepared in accordance with ISO 8655-5, ISO 8655-6 and ISO/TR 20461 documents • in Laboratory calibration is done. Declared uncertainty are the value at 100% of the nominal volume.
Volumetric Glassware Volumetric Flask	1 mL \leq V \leq 10 mL	Filling Glassware	8 μ L	V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	20 mL	Filling Glassware	12 μ L	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Volumetric Flask	25 mL	Filling Glassware	12 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	50 mL	Filling Glassware	18 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	100 mL	Filling Glassware	28 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	200 mL	Filling Glassware	40 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	250 mL	Filling Glassware	40 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	500 mL	Filling Glassware	70 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Volumetric Flask	1000 mL	Filling Glassware	0,10 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	2000 mL	Filling Glassware	0,16 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	5000 mL	Filling Glassware	0,34 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 1042, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	10 mL	Filling Plasticware	12 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	25 mL	Filling Plasticware	12 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	50 mL	Filling Plasticware	18 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Volumetric Flask	100 mL	Filling Plasticware	26 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	250 mL	Filling Plasticware	42 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	500 mL	Filling Plasticware	70 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Volumetric Flask	1000 mL	Filling Plasticware	0,12 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 5215, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	5 mL	Filling Glassware	16 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	10 mL	Filling Glassware	30 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Graduated Cylinder -	25 mL	Filling Glassware	62 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	50 mL	Filling Glassware	0,12 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	100 mL	Filling Glassware	0,14 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	250 mL	Filling Glassware	0,26 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	500 mL	Filling Glassware	0,60 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	1000 mL	Filling Glassware	1,2 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Graduated Cylinder -	2000 mL	Filling Glassware	2,4 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 4788, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	10 mL	Filling Plasticware	32 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	25 mL	Filling Plasticware	0,16 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	50 mL	Filling Plasticware	0,30 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	100 mL	Filling Plasticware	0,32 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	250 mL	Filling Plasticware	0,64 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Graduated Cylinder -	500 mL	Filling Plasticware	1,6 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	1000 mL	Filling Plasticware	3,2 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	2000 mL	Filling Plasticware	6,6 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Graduated Cylinder -	4000 mL	Filling Plasticware	16 mL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 6706, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	0,1 mL ≤ V ≤ 0,5 mL	Emptying Glassware	1,8 µL	V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	1 mL	Emptying Glassware	2,2 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Pipette (Graduated)	2 mL	Emptying Glassware	3,0 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	5 mL	Emptying Glassware	9,0 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	10 mL	Emptying Glassware	15 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	20 mL	Emptying Glassware	30 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Graduated)	25 mL	Emptying Glassware	30 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 835, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	0,5 mL	Emptying Glassware	1,4 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Pipette (Monosize)	1 mL	Emptying Glassware	2,4 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	2 mL	Emptying Glassware	3,2 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	5 mL	Emptying Glassware	4,2 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	10 mL	Emptying Glassware	6,0 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	25 mL	Emptying Glassware	8,0 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	50 mL	Emptying Glassware	12 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Pipette (Monosize)	100 mL	Emptying Glassware	20 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pipette (Monosize)	200 mL	Emptying Glassware	30 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 648, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	1 mL	Emptying Glassware	1,8 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	2 mL	Emptying Glassware	3,2 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	5 mL	Emptying Glassware	3,2 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	10 mL	Emptying Glassware	6,0 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Burette	25 mL	Emptying Glassware	10 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	50 mL	Emptying Glassware	16 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Burette	100 mL	Emptying Glassware	32 µL	The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 385, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pycnometer	1 mL ≤ V ≤ 100 mL	Gay-Lussac	7,6 µL	V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 3507, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pycnometer	10 mL ≤ V ≤ 100 mL	Reischauer	4,4 µL	V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 3507, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.
Volumetric Glassware Pycnometer	25 mL ≤ V ≤ 50 mL	Hubbard	32 µL	V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 3507, ISO 384, ISO 4787 and EURAMET cg.19 documents • in Laboratory calibration is done.

Accreditation Scope




UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Volumetric Glassware Pycnometer	10 mL ≤ V ≤ 100 mL	Ground-in thermometer	10 µL	<p>V: Measured Volume The values given in the measuring range column are the nominal volume of the device. According to the calibration procedure prepared in accordance with ISO 3507, ISO 384, ISO 4787 and EURAMET cg.19 documents</p> <ul style="list-style-type: none"> in Laboratory calibration is done.
Volumetric Glassware Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks	1 L	Filling or Discharging	0,35 mL	<p>The values given in the measuring range column are the nominal volume of the device.</p> <p>Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Volumetric Glassware Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks	2 L	Filling or Discharging	0,62 mL	<p>The values given in the measuring range column are the nominal volume of the device.</p> <p>Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Volumetric Glassware Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks	5 L	Filling or Discharging	1,03 mL	<p>The values given in the measuring range column are the nominal volume of the device.</p> <p>Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Volumetric Glassware Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks	10 L	Filling or Discharging	1,3 mL	<p>The values given in the measuring range column are the nominal volume of the device.</p> <p>Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site
Volumetric Glassware Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks	20 L	Filling or Discharging	2,3 mL	<p>The values given in the measuring range column are the nominal volume of the device.</p> <p>Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document.</p> <ul style="list-style-type: none"> Carried out in laboratory or in customer's site

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>		UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ		
		Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026		
Volumetric Glassware	50 L	Filling or Discharging	5,1 mL	The values given in the measuring range column are the nominal volume of the device. Calibration is performed by (*) Volumetric method according to the calibration procedure prepared in accordance with EURAMET cg.21 document. • Carried out in laboratory or in customer's site
Standard Capacity Measures, Volumetric Scales,, Liter Caps,, Fuel Mobile Scales,, Large Proving Tanks				

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p>
<p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>	

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Torque

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
<p>Torque Measurement Devices Torque Converter (Torque Sensor / Torque Calibrator)</p>	<p>$1 \text{ N}\cdot\text{m} \leq M \leq 1000 \text{ N}\cdot\text{m}$</p>	<p>1st class and coarser torque sensors with Reference arms and dead weights (Clockwise and anticlockwise)</p>	<p>%0,2</p>	<p><i>M</i> : Measured Torque (N.m) Calibration is performed according to the calibration procedure prepared in accordance with the EURAMET cg.14 document. • Carried out in laboratory or in customer's site</p>
<p>Torque Measurement Devices Torque Multiplier</p>	<p>$100 \text{ N}\cdot\text{m} \leq M \leq 2000 \text{ N}\cdot\text{m}$</p>	<p>0.5 Class Load Cell with Reference Moment Arms (Clockwise)</p>	<p>%0,8</p>	<p><i>M</i> : Measured Torque (N.m) In accordance with the calibration procedure defined in the in-house method (SOP 3-9-05 Rev.00) document. • Carried out in laboratory</p>
<p>Torque Measurement Devices Lid/Cap Opening-Closing Torque Device Torque Sensor</p>	<p>$1 \text{ N}\cdot\text{m} \leq M \leq 50 \text{ N}\cdot\text{m}$</p>	<p>Reference Moment Arm and With Dead Weight</p>	<p>%0,4</p>	<p><i>M</i> : Measured Torque (N.m) In accordance with the calibration procedure defined in the in-house method (SOP 3-9-03 Rev.00) document. • Carried out in laboratory</p>
<p>Torque Measurement Devices Torque Hand Tools</p>	<p>$0,2 \text{ N}\cdot\text{m} \leq M \leq 813 \text{ N}\cdot\text{m}$</p>	<p>With 1st Class Torque Sensor (Clockwise and anticlockwise)</p>	<p>%0,6</p>	<p><i>M</i> : Measured Torque (N.m) Calibration is performed according to the calibration procedure prepared in accordance with the ISO 6789-2 document. • Carried out in laboratory</p>
<p>Torque Measurement Devices Torque Hand Tools</p>	<p>$100 \text{ N}\cdot\text{m} \leq M \leq 2000 \text{ N}\cdot\text{m}$</p>	<p>0.5 Class Load Cell with (Clockwise and anticlockwise)</p>	<p>%0,7</p>	<p><i>M</i> : Measured Torque (N.m) Calibration is performed according to the calibration procedure prepared in accordance with the ISO 6789-2 document. • Carried out in laboratory</p>
<p>Torque Measurement Devices Continuous rotation torque tool Rotary tools for threaded fasteners Electric torque tool Pneumatic torque tool Hydraulic torque tool Torque wrench</p>	<p>$0,5 \text{ N}\cdot\text{m} \leq M \leq 450 \text{ N}\cdot\text{m}$</p>	<p>With 1st Class Torque Sensor (Clockwise and anticlockwise)</p>	<p>% 1,3</p>	<p><i>M</i> : Measured Torque (N.m) In accordance with the calibration procedure defined in the in-house method (SOP 3-9-04 Rev.00) document. • Carried out in laboratory • Carried out at the customer's site</p>

This document has been signed by Gülден Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Density

Ölçüm Büyüklüğü / Kalibre Edilen Cihazlar	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Hydrometer (Density measuring device) Specific Gravity (Relative Density) Hydrometers	$600 \text{ kg/m}^3 \leq \rho \leq 2000 \text{ kg/m}^3$	Hydrostatics Weighing Management (Cuckow)	0,4 kg/m ³	Calibration is performed according to the calibration procedure prepared in accordance with the NIST SP 250-78 document. <ul style="list-style-type: none">Carried out in laboratory
Hydrometer (Density measuring device) Bomehydrometer	0-70 °Be	Hydrostatics Weighing Management (Cuckow)	0,2 °Be	Calibration is performed according to the calibration procedure prepared in accordance with the NIST SP 250-78 document. <ul style="list-style-type: none">Carried out in laboratory
Hydrometer (Density measuring device) Alcoholimeter (% Hydrometers)	0-100 %	Hydrostatics Weighing Management (Cuckow)	0,025 %	Calibration is performed according to the calibration procedure prepared in accordance with the NIST SP 250-78 document. <ul style="list-style-type: none">Carried out in laboratory

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Mass (Mass Standards)

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Weight Standard E2 Standard Weights	1 mg ≤ m ≤ 20 mg	-	0,003 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	50 mg		0,004 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	100 mg		0,005 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	200 mg		0,006 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	500 mg		0,008 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	1 g		0,010 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	2 g		0,012 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	5 g		0,016 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	10 g		0,020 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard E2 Standard Weights	20 g		0,025 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	50 g		0,030 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	100 g		0,05 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	200 g		0,10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	500 g		0,25 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	1 kg		0,5 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	2 kg		1,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard E2 Standard Weights	5 kg		2,5 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	$1 \text{ mg} \leq m \leq 5 \text{ mg}$		0,006 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	10 mg		0,008 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	20 mg		0,010 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard F1 Standard Weights	50 mg		0,012 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	100 mg		0,016 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	200 mg		0,020 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	500 mg		0,025 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	1 g		0,03 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	2 g		0,04 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	5 g		0,05 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	10 g		0,06 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	20 g		0,08 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	50 g		0,10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	100 g		0,16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard F1 Standard Weights	200 g		0,30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	500 g		0,80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	1 kg		1,6 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	2 kg		3,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	5 kg		8,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	10 kg		16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	20 kg		30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F1 Standard Weights	50 kg		80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	$1 \text{ mg} \leq m \leq 5 \text{ mg}$		0,020 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	10 mg		0,025 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	20 mg		0,03 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard F2 Standard Weights	50 mg		0,04 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	100 mg		0,05 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	200 mg		0,06 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	500 mg		0,08 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	1 g		0,10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	2 g		0,12 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	5 g		0,16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	10 g		0,20 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	20 g		0,25 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	50 g		0,30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	100 g		0,50 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>			
---	---	--	--	--

Weight Standard F2 Standard Weights	200 g		1,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	500 g		2,5 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	1 kg		5,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	2 kg		10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	5 kg		25 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	10 kg		50 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	20 kg		100 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard F2 Standard Weights	50 kg		250 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	$1 \text{ mg} \leq m \leq 5 \text{ mg}$		0,06 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	10 mg		0,08 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	20 mg		0,10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard M1 Standard Weights	50 mg		0,12 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	100 mg		0,16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	200 mg		0,20 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	500 mg		0,25 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	1 g		0,30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	2 g		0,40 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	5 g		0,50 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	10 g		0,60 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	20 g		0,80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	50 g		1,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	100 g		1,6 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard M1 Standard Weights	200 g		3,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	500 g		8,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	1 kg		16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	2 kg		30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	5 kg		80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M1 Standard Weights	10 kg		160 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M1 Standard Weights	20 kg		300 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M1 Standard Weights	50 kg		800 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M2 Standard Weights	100 mg		0,50 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	200 mg		0,60 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	500 mg		0,80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	1 g		1,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard M2 Standard Weights	2 g		1,2 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	5 g		1,6 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	10 g		2,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	20 g		2,5 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	50 g		3,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	100 g		5,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	200 g		10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	500 g		25 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	1 kg		50 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	2 kg		100 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M2 Standard Weights	5 kg		250 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard M2 Standard Weights	10 kg		500 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M2 Standard Weights	20 kg		1000 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M2 Standard Weights	50 kg		2500 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	1 g		3,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	2 g		4,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	5 g		5,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	10 g		6,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	20 g		8,0 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	50 g		10 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	100 g		16 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	200 g		30 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory
Weight Standard M3 Standard Weights	500 g		80 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Weight Standard M3 Standard Weights	1 kg		160 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	2 kg		300 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	5 kg		800 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	10 kg		1600 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	20 kg		3000 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Weight Standard M3 Standard Weights	50 kg		8000 mg	Calibration is performed by determining the conventional mass value with the calibration procedure prepared in accordance with the OIML R-111 document. • Carried out in laboratory or in customer's site
Non-Standard Mass	$1 \text{ g} \leq m < 1 \text{ kg}$		1,8 mg	Calibration is performed by determining the conventional mass value by weighing on the balance with reference to the F1 class mass. • Carried out in laboratory
Non-Standard Mass	$1 \text{ kg} \leq m \leq 5 \text{ kg}$		4,0 mg	Calibration is performed by determining the conventional mass value by weighing on the balance with reference to the F1 class mass. • Carried out in laboratory
Non-Standard Mass	$5 \text{ kg} < m \leq 20 \text{ kg}$		30 mg	Calibration is performed by determining the conventional mass value by weighing on the balance with reference to the F1 class mass. • Carried out in laboratory
Non-Standard Mass	$20 \text{ kg} < m \leq 51 \text{ kg}$		100 mg	Calibration is performed by determining the conventional mass value by weighing on the balance with reference to the F1 class mass. • Carried out in laboratory

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Material Testing Machines

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Material Testing Machines Tensile / Compression Testing Machine Compression Testing Machine Tensile Testing Machine	$0,01 \text{ N} \leq F \leq 1000 \text{ N}$	Dead Weight Pulling Direction	%0,10	Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 7500-1, ISO 7500-1 and ASTM E4 documents. • Carried out at the customer's site
Material Testing Machines Tensile / Compression Testing Machine Compression Testing Machine Tensile Testing Machine	$100 \text{ N} \leq F \leq 100 \text{ kN}$	With 0.5 Class Load Cell Tensile Direction	%0,16	Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 7500-1, ISO 7500-1 and ASTM E4 documents. • Carried out at the customer's site
Material Testing Machines Tensile / Compression Testing Machine Compression Testing Machine Tensile Testing Machine	$20 \text{ kN} \leq F \leq 500 \text{ kN}$	Tension in compression with cage	%0,16	Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 7500-1, ISO 7500-1 and ASTM E4 documents. • Carried out at the customer's site
Material Testing Machines Tensile / Compression Testing Machine Compression Testing Machine Tensile Testing Machine	$10 \text{ N} \leq F \leq 500 \text{ kN}$	0.5 class load cell in compression direction	%0,16	Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 7500-1, ISO 7500-1 and ASTM E4 documents. • Carried out at the customer's site
Material Testing Machines Tensile / Compression Testing Machine Compression Testing Machine Tensile Testing Machine	$200 \text{ kN} \leq F \leq 3 \text{ MN}$	1 class load cell in compression direction	%0,32	Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 7500-1, ISO 7500-1 and ASTM E4 documents. • Carried out at the customer's site
Extensometer	$0,5 \text{ mm} \leq L \leq 1000 \text{ mm}$	With Laser Interferometer	$(0,7 + 1,5.L) \mu\text{m}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 9513 ve ASTM E83 documents. • Carried out at the customer's site
Concrete Test Presses	$200 \text{ kN} \leq F \leq 3 \text{ MN}$	With 1st Class Load Cell	%0,32	Calibration is performed according to the calibration procedure prepared in accordance with TS EN 12390-4 documents. • Carried out at the customer's site

Accreditation Scope



UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ

Accreditation Nr : AB-0012-K
Revision Nr: 23 Date: 15.04.2026

Notch-Impact Tester Izod Impact Tester	$0,5 \text{ J} \leq A_p \leq 750 \text{ J}$	Force Pendulum Length Angle Time	%0,16 0,3 mm 0,1° 0,03 s	Measurement uncertainty for the following parameters is calculated: 1-Centre of impact 2-Potential energy 3-Indicator deviation Calibration is performed according to the calibration procedure prepared in accordance with TS EN ISO 148-2, ASTM E23 ve TS EN ISO 13802 documents. • Carried out at the customer's site
Spring Impact Test Hammer	$0,1 \text{ J} \leq A_p \leq 0,3 \text{ J}$	with spring impact test hammer calibrator	0,01 J	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN 60068-2-75 document. • Carried out in laboratory
Spring Impact Test Hammer	$0,3 \text{ J} < A_p \leq 2,0 \text{ J}$	with spring impact test hammer calibrator	0,015 J	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN 60068-2-75 document. • Carried out in laboratory
Spring Impact Test Hammer Calibrator	$0,1 \text{ J} \leq A_p \leq 0,3 \text{ J}$	with spring impact test hammer calibrator calibration stand	0,007 J	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN 60068-2-75 document. • Carried out in laboratory
Spring Impact Test Hammer Calibrator	$0,3 \text{ J} < A_p \leq 2,0 \text{ J}$	with spring impact test hammer calibrator calibration stand	0,008 J	Calibration is performed according to the calibration procedure prepared in accordance with the TS EN 60068-2-75 document. • Carried out in laboratory
Material Testing Space in Machines Modification and Measurement Systems	$0,5 \text{ mm} \leq L \leq 1000 \text{ mm}$	With Laser Interferometer	$(0,7 + 1,5.L) \mu\text{m}$	L : Measured Value [m] Calibration is performed according to the calibration procedure prepared in accordance with ASTM E2309 documents. • Carried out at the customer's site
Material Testing Machines Feed Rate Calibration	$0,05 \text{ mm/min} \leq v \leq 500 \text{ mm/min}$	With Laser Interferometer Tensile-Compression Directions.	%0,17	Calibration is performed according to the calibration procedure prepared in accordance with ASTM E2658 documents. • Carried out at the customer's site

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 <p>Calibration TS EN ISO/IEC 17025 AB-0012-K</p>	<p>UMS ANKARA KALİBRASYON MÜHENDİSLİK MÜŞAVİRLİK MÜMESSİLLİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ</p> <p>Accreditation Nr : AB-0012-K Revision Nr: 23 Date: 15.04.2026</p>
--	---

Calibration and Measurement Capability (CMC)

Mechanical Quantities/Force

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
Force Measurement Devices Load Cell Force Converter Dynamometer Hand Scale Crane Scale	$0,01 N \leq F \leq 1000 N$	With standard masses (with dead weights) Tension / Compression direction	% 0,10	Calibration is performed according to the calibration procedure prepared in accordance with the DKD R 3-3 document. <ul style="list-style-type: none">Carried out in laboratory or in customer's site
Force Measurement Devices Load Cell Force Converter Dynamometer Hand Scale Crane Scale	$0,1 kN \leq F \leq 50 kN$	With 0.5 class load cell Tensile-Compression direction	% 0,16	Calibration is performed according to the calibration procedure prepared in accordance with the DKD R 3-3 document. <ul style="list-style-type: none">Carried out in laboratory or in customer's site
Force Measurement Devices Single Capacity Force Gauges	$1 N \leq F \leq 50 kN$	With 0.5 Class load cell	% 0,20	Calibration is performed using with loadcell the comparison method <ul style="list-style-type: none">Carried out in laboratory or in customer's site

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.